

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

AUG 24 2009

Supervisor/Operator Performing the Verification Procedure:

DataMaster cdm S/N 130666 ✓

Name Gary Yamamoto ID# 2765 Date 8-19-09 ✓

A

Agency Fairbanks Police Department Phone # 459-6500

Instrument Location 911 Cushman Street Fairbanks, Alaska 99701

Alco S/N X124627 ✓ Target Value .074 ✓ High Pressure 1300

B

Alco Test Value Average .076 ✓ .076 ✓
1st Alco 2nd Alco

Signature [Signature] ✓

CAB
8/31/09

(OVER)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows: (Do Not write in the area below)


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 17th day of Sept., 2009.

[Signature] (Notary Seal Stamp)

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT
of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N _____ 130666 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130666 ✓

AUGUST 19, 2009 ✓ *9/2/09*

OPERATOR'S NAME:

YAMAMOTO

OPERATOR'S NUMBER: 2765

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI:

YOC

O.L. #: 1111

DEPT/AGENCY: FBK1

CASE/REPORT: 1111

TEST TYPE: V

ALCO TARGET VALUE: .074 ✓

ALCO S/N: X124627 ✓

--- BREATH ANALYSIS ---

.074 ADJUSTED FOR 29.94 in		
ALCO TARGET	.074	17:24
BLANK TEST	.000	17:26
INTERNAL STANDARD	VERIFIED	17:26
ALCO TV 29.94 in	.076 ✓	17:26
BLANK TEST	.000	17:27
SUBJECT SAMPLE <i>9/2/09</i>	.000	17:27
BLANK TEST	.000	17:28
ALCO TV 29.94 in	.076 ✓	17:29
BLANK TEST	.000	17:29

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130666

AUGUST 19, 2009 ✓ *9/2/09*
TIME 17:31

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	43c
BREATH TUBE:	42c
BAROMETER:	29.94 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKLMNPOQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~"